

Lead/Pesticide Test


1 Open foil pouch marked "Pe/Pb." Using dropper, place exactly TWO droppers full of water sample into small test vial included in pouch. To pick up sample, tightly squeeze the bulb at the end of the dropper and place the open end into water sample. Release the bulb to pick up sample, then squeeze again to expel sample into vial.



2 Swirl vial gently for several seconds. Place on a flat surface.



3 Place both test strips into the test vial with arrows pointing DOWN.



4 Wait 10 minutes. Do not disturb strips or vial during this time. Blue lines will appear on the strips.




5 Take the strips out of the vial and lay them on a flat surface with the arrows pointing to the LEFT. Read results.



NEGATIVE RESULT (less than USEPA action level of 15 ppb): **LEFT** line, next to number 1, is darker than the **RIGHT** line, next to number 2 **OR** if you only see one line the result is negative.



POSITIVE RESULT (greater than USEPA action level of 15 ppb): **RIGHT** line, next to number 2, is darker than the **LEFT** line, next to number 1, **OR** if both **LEFT** and **RIGHT** lines are equally dark the result is positive.



Note: If no lines appear, or both lines are very light, the test did not run properly and the result is not valid.